

OCT 09 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 09 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130579</u> ✓	
Name <u>C. RIALA</u>		ID# <u>3791</u>	Date <u>10/8/09</u> ✓
A	Agency <u>APD</u>		Phone # <u>786-2640</u>
Instrument Location <u>Anch. Jail</u> ✓			
B			
Alco S/N <u>X301563</u> ✓		Target Value <u>.079</u> ✓	High Pressure <u>100</u>
Alco Test Value Average <u>083</u> ✓ 1 st Alco		<u>083</u> ✓ 2 nd Alco	
Signature <u>[Signature]</u>		<u>1349</u>	<u>008</u> <u>10/12/09</u>
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of Oct., 2009.

[Signature]
Carolyn M. Noland

Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130579 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

OCTOBER 08, 2009

OPERATOR'S NAME:

RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

A

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: V

ALCO TARGET VALUE: 079

ALCO S/N: X301563

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.20 in
ALCO TARGET .079 22:31
BLANK TEST .000 22:32
INTERNAL STANDARD VERIFIED 22:32
ALCO TV 30.20 in .083 22:32
BLANK TEST .000 22:33
SUBJECT SAMPLE .000 22:33
BLANK TEST .000 22:34
ALCO TV 30.20 in .083 22:35
BLANK TEST .000 22:35

CR/1349 10/8/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

OCTOBER 08, 2009
TIME 22:30

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/08
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.20 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¢

CR/1349

10/8/09